

# **ECAT FOUNDATION**

**External quality Control for Assays and Tests**

**With a focus on Thrombosis and Haemostasis**

## **REPORT**



**SURVEY 2025-M4**  
**Main**  
**Labcode 1492**

Date of Issue : 17-januari-2026

Survey : 2025-M4

Report : Main

**Note:**

In the Survey Manual 2025 detailed information is given regarding the ECAT external quality assessment programme, including the statistical evaluation and explanation of the report.

This Survey Manual 2025 should be considered as an integral part of this survey report.

Please notice the information regarding the homogeneity of samples used and the between-laboratory variation in the paragraph on the statistical evaluation of the Survey Manual.

**General Information****Exclusion of results**

Results < [value] or > [value] are excluded from the statistical analysis. When other results (e.g. deviating results) are excluded from the statistical analysis, these results are placed between brackets.

**Multiple instruments**

Participants have the option to submit three results per parameter for up to three different instruments and /or different methods. However, some participants submitted identical results three times, implying these were separate results. This is not the case and may affect the statistical analysis. **The results submitted for the second and third system have been excluded from the statistical analysis. Please submit results no more than once unless they are a different measurement.**

**Complaints**

Any complaints regarding this survey report should be reported to the ECAT before **March 10th, 2026**. Complaints received after this date will not be taken into consideration.

This report is authorized by: Dr. M.J. van Essen-Hollestelle (Programme Expert)

**Note:** The current Survey Manual can be downloaded from the 'Participant Area' of the ECAT website. Login and select the option 'View Documents'.

ECAT Foundation

Director: Dr. P. Meijer  
ECAT Office  
P.O. Box 107  
2250 AC Voorschoten, The Netherlands  
phone +31 (0) 71 3030 910; fax + 31 (0) 71 3030 919  
E-mail: [info@ecat.nl](mailto:info@ecat.nl)  
Website: [www.ecat.nl](http://www.ecat.nl)  
VAT number: NL802836872B01

Registration number with the Chamber of Commerce (KvK) Gouda : 41174102  
General terms of delivery are applicable to all our services.

All rights reserved. No part of this report may be reproduced, stored in a retrieval system, or by any means, electronic, mechanical, photocopying, recording or otherwise, without prior permission from the ECAT Foundation.

## OVERVIEW Z-SCORES

 UV:Satisfactory (-2 ≤ Z-score ≤ 2)

 UV: Need attention (-3 < Z-score < -2 or 2 < Z-score < 3)

 UV:Unsatisfactory (Z-score ≤ -3 or ≥ 3)  
 BV: Unacceptable

Module	Parameter	Sample	Test System	Total	Assay	Method	Bivariate Analysis
ADAMTS13 - I	ADAMTS13 Activity	25.220	1	-0.76	-0.69	-0.84	
		25.221	1	-0.80	-0.53	-0.74	

**ADAMTS13 - I**
**ADAMTS13 Activity**

Sample No	25.220		
Sample Details	Normal Coagulation Control Plasma		
Prior Use	None		
Unit	% or IU/dL		
Expiry Date	30-september-2026		
Homogeneity	0.8 %	<b>Homogeneity Parameter</b>	APTT
		For any method used for the measurement of this parameter with a <b>CV ≤ 2.7%</b> the criterion for homogeneity could not be met and the Z-scores should be interpreted with caution. See for further details the paragraph on the statistical evaluation in the Survey Manual.	
Number of Participants	205		
Number of Responders	190	<b>Response Rate</b>	93 %
Comments	The following participant reported a deviating result which was excluded from the statistical analysis: <b>9065:</b> 18%		

The result of participant (**labcode 610**) was excluded because they reported the result with an incorrect unit.

Participant (**labcode 513**) reported a deviating result which was excluded from the statistical analysis. This result might be caused by an accidental mix-up of the sample 25.220 and 25.221.

Classification Overview	Normal	BorderLine normal	Borderline abnormal	Abnormal	No classification
<b>Total</b>	213	1	0	1	3



Your Classification      Normal

**Own Reagent**

	n	assigned value	Uncert.	CV (%)	Range	your result	z-score	your result	z-score	your result	z-score
<b>Total Group</b>	209	101	1.14	13.0	45 - 145	91.1	-0.76				
<b>FRET</b>	59	98	1.70	10.6	45 - 131	91.1	-0.69				
Technoclone Technofluor activity	39	98	1.63	8.3	79 - 115	91.1	-0.84				

Other Reagents

	n	assigned value	Uncert.	CV (%)	Range
<b>Total Group</b>	209	101	1.14	13.0	45 - 145
<b>FRET</b>	59	98	1.70	10.6	45 - 131
<b>Homemade</b>	13	99	4.69	13.6	75 - 117
<b>Immuclor activity</b>	2	126			120 - 131
<b>Other</b>	4	83			45 - 108
<b>Sekisui Actifluor activity</b>	1	91			-
<b>Enzyme Immuno Assay</b>	57	99	2.47	15.1	64 - 134
<b>Technoclone Technozym activity</b>	57	99	2.47	15.1	64 - 134
<b>Latex Immuno Assay</b>	1	87			-
<b>Homemade</b>	1	87			-
<b>CLIA/FIA</b>	89	105	1.78	12.8	66 - 145
<b>Werfen Acustar activity</b>	89	105	1.78	12.8	66 - 145
<b>Other</b>	3	92			75 - 113
<b>Homemade</b>	2	94			75 - 113
<b>Other</b>	1	92			-



**ADAMTS13 - I**
**ADAMTS13 Activity**

<b>Sample No</b>	25.221		
<b>Sample Details</b>	Plasma with an ADAMTS13 activity level between 15 and 25%		
<b>Prior Use</b>	None		
<b>Unit</b>	% or IU/dL		
<b>Expiry Date</b>	31-juli-2027		
<b>Homogeneity</b>	4.2 %	<b>Homogeneity Parameter</b>	ADAMTS13
		For any method used for the measurement of this parameter with a <b>CV ≤ 14.0%</b> the criterion for homogeneity could not be met and the Z-scores should be interpreted with caution. See for further details the paragraph on the statistical evaluation in the Survey Manual.	
<b>Number of Participants</b>	205		
<b>Number of Responders</b>	190	<b>Response Rate</b>	93 %
<b>Comments</b>	The following participants reported deviating results which were excluded from the statistical analysis:		
724:	0 %		
9005:	2 %		
9074:	1%		

The result of participant (**labcode 610**) was excluded because they reported the result with an incorrect unit.

Participant (**labcode 513**) reported a deviating result which was excluded from the statistical analysis. This result might be caused by an accidental mix-up of the sample 25.220 and 25.221.

Classification Overview	Normal	BorderLine normal	Borderline abnormal	Abnormal	No classification
<b>Total</b>	8	1	11	194	4



Your Classification      Abnormal

**Own Reagent**










	n	assigned value	Uncert.	CV (%)	Range	your result	z-score	your result	z-score	your result	z-score
<b>Total Group</b>	214	22	0.29	15.6	6 - 55	19	-0.80				
<b>FRET</b>	59	21	0.56	16.5	7 - 30	19	-0.53				
Technoclone Technofluor activity	39	21	0.52	12.5	9 - 25	19	-0.74				

Other Reagents

	n	assigned value	Uncert.	CV (%)	Range
<b>Total Group</b>	214	22	0.29	15.6	6 - 55
<b>FRET</b>	59	21	0.56	16.5	7 - 30
<b>Homemade</b>	13	20	2.19	31.5	7 - 30
<b>Immucor activity</b>	3	25			23 - 30
<b>Other</b>	4	15			9 - 24
<b>Enzyme Immuno Assay</b>	62	24	0.67	17.7	6 - 55
<b>Technoclone Technozym activity</b>	62	24	0.67	17.7	6 - 55
<b>Latex Immuno Assay</b>	1	16			-
<b>Homemade</b>	1	16			-
<b>CLIA/FIA</b>	89	21	0.28	10.0	16 - 34
<b>Werfen Acustar activity</b>	89	21	0.28	10.0	16 - 34
<b>Other</b>	3	23			18 - 28
<b>Homemade</b>	2	20			18 - 23
<b>Other</b>	1	28			-



**ADAMTS13 - I**
**ADAMTS13 Activity**

 Instrument ID 

**Z-score history (by survey)**

**Z-score history (by consensus value)**

 Instrument ID 

**Z-score history (by survey)**

**Z-score history (by consensus value)**

 Instrument ID 

**Z-score history (by survey)**

**Z-score history (by consensus value)**




**ADAMTS13 - I**
**ADAMTS13 antigen**

Sample No	25.220
Sample Details	Normal Coagulation Control Plasma
Prior Use	None
Unit	IU/mL
Expiry Date	30-september-2026
Homogeneity	0.8 %
	<b>Homogeneity Parameter</b>
	APTT
	For any method used for the measurement of this parameter with a <b>CV <math>\leq</math> 2.7%</b> the criterion for homogeneity could not be met and the Z-scores should be interpreted with caution. See for further details the paragraph on the statistical evaluation in the Survey Manual.
Number of Participants	205
Number of Responders	23
Comments	The result of participant ( <b>labcode 950</b> ) was excluded because they reported the result with an incorrect unit.

Classification Overview	Normal	BorderLine normal	Borderline abnormal	Abnormal	No classification
<b>Total</b>	23	0	1	0	1

Other Reagents

	n	assigned value	Uncert.	CV (%)	Range
<b>Enzyme Immuno Assay</b>	24	0.83	0.05	24.7	0.41 - 1.39
Homemade	1	1.32			-
Other	1	0.75			-
Sekisui Imubind antigen	1	1.17			-
Technozym Technozym activity/antigen	2	0.94			0.83 - 1.06
Technozym Technozym antigen	19	0.78	0.05	21.5	0.41 - 1.39



□ Overall



**ADAMTS13 - I**
**ADAMTS13 antigen**

Sample No	25.221		
Sample Details	Plasma with an ADAMTS13 activity level between 15 and 25%		
Prior Use	None		
Unit	IU/mL		
Expiry Date	31-juli-2027		
Homogeneity	4.2 %	<b>Homogeneity Parameter</b>	ADAMTS13
		For any method used for the measurement of this parameter with a <b>CV ≤ 14.0%</b> the criterion for homogeneity could not be met and the Z-scores should be interpreted with caution. See for further details the paragraph on the statistical evaluation in the Survey Manual.	
Number of Participants	205		
Number of Responders	23	<b>Response Rate</b>	11 %
Comments	The result of participant ( <b>labcode 950</b> ) was excluded because they reported the result with an incorrect unit.		

Classification Overview	Normal	BorderLine normal	Borderline abnormal	Abnormal	No classification
<b>Total</b>	3	3	2	16	1

Other Reagents

	n	assigned value	Uncert.	CV (%)	Range
<b>Enzyme Immuno Assay</b>	24	0.34	0.02	19.4	0.19 - 0.75
Homemade	1	0.39			-
Other	1	0.19			-
Sekisui Imubind antigen	1	0.28			-
Technozym Technozym activity/antigen	2	0.35			0.22 - 0.49
Technozym Technozym antigen	19	0.34	0.01	12.9	0.28 - 0.75




Overall


**ADAMTS13 - I**
**ADAMTS13 antigen**

Instrument ID	
---------------	--


• Needs attention    ■ Unsatisfactory    ▲ Outlier    ● Your Result



**Z-score history (by survey)**




**Z-score history (by consensus value)**




Instrument ID	
---------------	--


• Needs attention    ■ Unsatisfactory    ▲ Outlier    ● Your Result



**Z-score history (by survey)**




**Z-score history (by consensus value)**




Instrument ID	
---------------	--

• Needs attention    ■ Unsatisfactory    ▲ Outlier    ● Your Result



**Z-score history (by survey)**



**Z-score history (by consensus value)**

